

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031919\
 Data File : VN054432.D
 Acq On : 19 Mar 2019 11:17
 Operator : JC/SP
 Sample : VN0319WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0319WBL01

Quant Time: Mar 20 04:04:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	2015149	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2850880	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2374367	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	903820	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	1027329	48.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.00%	
35) Dibromofluoromethane	7.59	113	971001	52.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.62%	
50) Toluene-d8	10.09	98	3723972	54.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.68%	
62) 4-Bromofluorobenzene	12.40	95	1121239	47.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.38%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031919\
Data File : VN054432.D
Acq On : 19 Mar 2019 11:17
Operator : JC/SP
Sample : VN0319WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0319WBL01

Quant Time: Mar 20 04:04:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

