

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031919\
 Data File : VN054433.D
 Acq On : 19 Mar 2019 11:47
 Operator : JC/SP
 Sample : K1970-05
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P-5-E.WALL-SOUTH

Quant Time: Mar 19 12:13:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1801786	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	2522587	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2235642	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1104404	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	782403	40.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.78%	
35) Dibromofluoromethane	7.59	113	724910	44.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.12%	
50) Toluene-d8	10.09	98	3037860	50.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.12%	
62) 4-Bromofluorobenzene	12.40	95	1000031	47.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.14%	
Target Compounds						
39) Methylcyclohexane	9.08	83	433344	15.161	ug/l	99
67) Ethyl Benzene	11.51	91	298283	3.666	ug/l	98
73) Isopropylbenzene	12.25	105	998655	11.302	ug/l	100
78) n-propylbenzene	12.59	91	2335113	24.356	ug/l	100
85) sec-Butylbenzene	13.17	105	2743745	31.274	ug/l	90
89) n-Butylbenzene	13.61	91	2386580	40.810	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031919\
Data File : VN054433.D
Acq On : 19 Mar 2019 11:47
Operator : JC/SP
Sample : K1970-05
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P-5-E.WALL-SOUTH

Quant Time: Mar 19 12:13:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

