

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031919\
 Data File : VN054437.D
 Acq On : 19 Mar 2019 13:36
 Operator : JC/SP
 Sample : K1930-10
 Misc : 5.82g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Quant Time: Mar 20 04:18:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1870781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	2661965	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2286263	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	988111	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	801905	40.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.72%	
35) Dibromofluoromethane	7.59	113	752868	43.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.70%	
50) Toluene-d8	10.09	98	3092828	48.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.56%	
62) 4-Bromofluorobenzene	12.40	95	970290	43.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.46%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031919\
Data File : VN054437.D
Acq On : 19 Mar 2019 13:36
Operator : JC/SP
Sample : K1930-10
Misc : 5.82g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Mar 20 04:18:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

