

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031919\
 Data File : VN054440.D
 Acq On : 19 Mar 2019 14:58
 Operator : JC/SP
 Sample : K1960-04DL 200X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-MW-9-9.0-031419DL

Quant Time: Mar 20 12:25:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	2069726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2994090	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2453065	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	971900	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	911827	41.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.96%	
35) Dibromofluoromethane	7.59	113	882720	45.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.42%	
50) Toluene-d8	10.09	98	3385489	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
62) 4-Bromofluorobenzene	12.40	95	1026388	41.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.26%	
Target Compounds						
16) Acetone	3.82	43	3437788	546.241	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031919\
Data File : VN054440.D
Acq On : 19 Mar 2019 14:58
Operator : JC/SP
Sample : K1960-04DL 200X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-MW-9-9.0-031419DL

Quant Time: Mar 20 12:25:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

