

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031919\
 Data File : VN054450.D
 Acq On : 19 Mar 2019 19:45
 Operator : JC/SP
 Sample : K1693-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-031819-B

Quant Time: Mar 20 04:56:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	2102640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2969706	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2495203	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1018362	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	897753	40.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.40%	
35) Dibromofluoromethane	7.59	113	874783	45.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.34%	
50) Toluene-d8	10.09	98	3409278	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
62) 4-Bromofluorobenzene	12.40	95	1066735	43.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.20%	
Target Compounds						
16) Acetone	3.82	43	23241	3.635	ug/l	95
20) Methylene Chloride	4.55	84	468994	22.340	ug/l	95
43) Isopropyl Acetate	8.17	43	31017	1.125	ug/l #	95
95) Naphthalene	15.13	128	274384	9.186	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031919\
Data File : VN054450.D
Acq On : 19 Mar 2019 19:45
Operator : JC/SP
Sample : K1693-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-03-031819-B

Quant Time: Mar 20 04:56:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

