

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031919\
 Data File : VN054451.D
 Acq On : 19 Mar 2019 20:12
 Operator : JC/SP
 Sample : K1941-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP

Quant Time: Mar 20 04:59:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	2155813	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	3074365	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2496680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	936397	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	929349	40.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.18%	
35) Dibromofluoromethane	7.59	113	901134	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.90%	
50) Toluene-d8	10.09	98	3478277	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
62) 4-Bromofluorobenzene	12.40	95	1039885	40.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.16%	
Target Compounds						
16) Acetone	3.82	43	18196	2.776	ug/l	99
20) Methylene Chloride	4.55	84	941772	43.753	ug/l	96
43) Isopropyl Acetate	8.17	43	30467	1.067	ug/l #	93
95) Naphthalene	15.13	128	55845	4.735	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031919\
Data File : VN054451.D
Acq On : 19 Mar 2019 20:12
Operator : JC/SP
Sample : K1941-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP

Quant Time: Mar 20 04:59:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

