

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031920\
 Data File : VN060630.D
 Acq On : 19 Mar 2020 12:44
 Operator : JC/MD
 Sample : PB127661ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127661ZHE#02

Quant Time: Mar 20 07:08:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	201422	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	335305	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	316740	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142068	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130429	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
35) Dibromofluoromethane	7.57	113	99727	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
50) Toluene-d8	10.08	98	415920	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	12.40	95	155344	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
Target Compounds						
16) Acetone	3.79	43	5883	6.006	ug/l	100
20) Methylene Chloride	4.52	84	6947	2.883	ug/l	99
43) Isopropyl Acetate	8.14	43	65927	11.395	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031920\
Data File : VN060630.D
Acq On : 19 Mar 2020 12:44
Operator : JC/MD
Sample : PB127661ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127661ZHE#02

Quant Time: Mar 20 07:08:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

