

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031920\
 Data File : VN060633.D
 Acq On : 19 Mar 2020 13:52
 Operator : JC/MD
 Sample : PB127661ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127661ZHE#05

Quant Time: Mar 20 03:55:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	205453	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	340156	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	324618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145371	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	134158	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
35) Dibromofluoromethane	7.56	113	99719	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
50) Toluene-d8	10.08	98	422002	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
62) 4-Bromofluorobenzene	12.40	95	158181	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
Target Compounds						
16) Acetone	3.79	43	6269	6.275	ug/l	98
20) Methylene Chloride	4.53	84	6579	2.677	ug/l	91
43) Isopropyl Acetate	8.14	43	62426	10.636	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031920\
Data File : VN060633.D
Acq On : 19 Mar 2020 13:52
Operator : JC/MD
Sample : PB127661ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127661ZHE#05

Quant Time: Mar 20 03:55:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

