

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031920\
 Data File : VN060634.D
 Acq On : 19 Mar 2020 14:15
 Operator : JC/MD
 Sample : PB127661ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127661ZHE#06

Quant Time: Mar 20 07:26:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	205355	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	341517	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	321750	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145505	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	132673	47.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.12%	
35) Dibromofluoromethane	7.56	113	100273	48.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.12%	
50) Toluene-d8	10.08	98	426884	50.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.22%	
62) 4-Bromofluorobenzene	12.40	95	158448	51.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.52%	
Target Compounds						
20) Methylene Chloride	4.52	84	4471	1.820	ug/l	93
43) Isopropyl Acetate	8.14	43	55610	9.437	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031920\
 Data File : VN060634.D
 Acq On : 19 Mar 2020 14:15
 Operator : JC/MD
 Sample : PB127661ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127661ZHE#06

Quant Time: Mar 20 07:26:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

