

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031920\
 Data File : VN060632.D
 Acq On : 19 Mar 2020 13:29
 Operator : JC/MD
 Sample : PB127661ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127661ZHE#04

Quant Time: Mar 20 03:53:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	205870	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	341151	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	321881	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	144137	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	133129	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
35) Dibromofluoromethane	7.57	113	101755	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
50) Toluene-d8	10.08	98	422823	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
62) 4-Bromofluorobenzene	12.40	95	158395	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
Target Compounds						
16) Acetone	3.79	43	6587	6.580	ug/l	93
20) Methylene Chloride	4.52	84	6997	2.841	ug/l	98
43) Isopropyl Acetate	8.14	43	63513	10.789	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031920\
Data File : VN060632.D
Acq On : 19 Mar 2020 13:29
Operator : JC/MD
Sample : PB127661ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127661ZHE#04

Quant Time: Mar 20 03:53:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1825

Delta R.T. -0.00 min

Lab File: VN060632.D

Acq: 19 Mar 2020 13:29

Instrument :

MSVOA_N

ClientSampleId :


PB127661ZHE#04

Tot Ion:168 Resp: 205870

Ion Ratio Lower Upper

168 100

99 57.4 44.0 66.0



Abundance Ion 167.90 (167.60 to 168.60): VN060565.D (-1807) (-)

Ion 98.90 (98.60 to 99.60): VN060565.D (-1807) (-)

7.64

80000

60000



40000

20000

0

7.60 7.70

Time-->



#16

Acetone

Concen: 6.580 ug/l

RT: 3.79 min Scan# 627

Delta R.T. 0.01 min

Lab File: VN060632.D


Acq: 19 Mar 2020 13:29

Tgt Ion: 43 Resp: 6587

Ion Ratio Lower Upper

43 100

58 27.3 25.1 37.7



Abundance Ion 43.00 (42.70 to 43.70): VN060565.D (-597) (-)

Ion 58.00 (57.70 to 58.70): VN060565.D (-597) (-)

3.79

2500

2000

1500


1000


500

0

3.70 3.75 3.80 3.85 3.90

Time-->









#20
Methvlene Chloride
Concen: 2.841 ug/l
RT: 4.52 min Scan# 853
Delta R.T. -0.01 min
Lab File: VN060632.D
Acq: 19 Mar 2020 13:29




Instrument :
MSVOA_N
ClientSampleId :
PB127661ZHE#04

Tot Ion:	84	Resp:	6997
Ion	Ratio	Lower	Upper
84	100		
49	126.4	103.3	154.9
51	39.4	31.6	47.4
86	64.1	52.0	78.0



#33
1,2-Dichloroethane-d4
Concen: 47.107 ug/l
RT: 8.00 min Scan# 1937
Delta R.T. -0.00 min
Lab File: VN060632.D
Acq: 19 Mar 2020 13:29

Tgt Ion:	65	Resp:	133129
Ion	Ratio	Lower	Upper
65	100		
67	53.4	0.0	107.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.57 min Scan# 2113

Delta R.T. -0.00 min

Lab File: VN060632.D

Acq: 19 Mar 2020 13:29

Instrument :

MSVOA_N

ClientSampleId :

PB127661ZHE#04


Tot Ion:114 Resp: 341151

Ion Ratio Lower Upper

114 100

63 22.4 0.0 43.8



88 16.2 0.0 33.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 49.328 ug/l

RT: 7.57 min Scan# 1802

Delta R.T. 0.00 min

Lab File: VN060632.D

Acq: 19 Mar 2020 13:29


Tgt Ion:113 Resp: 101755

Ion Ratio Lower Upper

113 100

111 102.5 82.9 124.3



192 19.9 15.8 23.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 10.789 ug/l

RT: 8.14 min Scan# 1981

Delta R.T. 0.00 min

Lab File: VN060632.D

Acq: 19 Mar 2020 13:29

Instrument :

MSVOA_N

ClientSampleId :

PB127661ZHE#04


Tot Ion: 43 Resp: 63513

Ion Ratio Lower Upper

43 100

61 19.5 20.7 31.1#




87 11.9 9.8 14.8



Abundance Ion 43.00 (42.70 to 43.70): VN060632.D (-1887) (-)

Ion 61.00 (60.70 to 61.70): VN060632.D (-1887) (-)

Ion 87.00 (86.70 to 87.70): VN060632.D (-1887) (-)



#50

Toluene-d8

Concen: 50.182 ug/l

RT: 10.08 min Scan# 2583

Delta R.T. -0.00 min

Lab File: VN060632.D


Acq: 19 Mar 2020 13:29

Tgt Ion: 98 Resp: 422823

Ion Ratio Lower Upper



98 100

100 65.0 51.8 77.8



Abundance Ion 98.00 (97.70 to 98.70): VN060632.D (-2490) (-)

Ion 100.00 (99.70 to 100.70): VN060632.D (-2490) (-)





#62

4-Bromofluorobenzene

Concen: 51.294 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. -0.00 min

Lab File: VN060632.D

Acq: 19 Mar 2020 13:29

Instrument :

MSVOA_N

ClientSampleId :

PB127661ZHE#04


Tot Ion: 95 Resp: 158395

Ion Ratio Lower Upper

95 100

174 78.4 0.0 157.6




176 75.3 0.0 150.6



Abundance Ion 94.90 (94.60 to 95.60): VN060632.D (-3211) (-)

Ion 173.80 (173.50 to 174.50): VN060632.D (-3211) (-)

Ion 175.80 (175.50 to 176.50): VN060632.D (-3211) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.40 min Scan# 2994

Delta R.T. -0.00 min

Lab File: VN060632.D

Acq: 19 Mar 2020 13:29


Tgt Ion: 117 Resp: 321881

Ion Ratio Lower Upper

117 100

82 56.6 46.7 70.1



119 32.1 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): VN060632.D (-2901) (-)

Ion 82.00 (81.70 to 82.70): VN060632.D (-2901) (-)

Ion 118.90 (118.60 to 119.60): VN060632.D (-2901) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. -0.00 min

Lab File: VN060632.D

Acq: 19 Mar 2020 13:29

Instrument :

MSVOA_N

ClientSampleId :

PB127661ZHE#04


Tot Ion:152 Resp: 144137

Ion Ratio Lower Upper

152 100

115 59.8 30.2 90.6

150 158.1 0.0 350.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

