

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031921\
 Data File : VN066283.D
 Acq On : 19 Mar 2021 16:15
 Operator : JC/MD
 Sample : M1707-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ELEVATOR-511-EL-162-157

Quant Time: Mar 20 04:59:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	287816	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	430817	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	397290	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	132688	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	200587	48.37	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.74%
35) Dibromofluoromethane	7.507	113	159893	58.18	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.36%
50) Toluene-d8	10.028	98	590173	55.54	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	111.08%
62) 4-Bromofluorobenzene	12.345	95	164140	42.85	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.70%
Target Compounds						
16) Acetone	3.741	43	89122	44.74	ug/l	100
20) Methylene Chloride	4.455	84	19988	6.41	ug/l #	82
43) Isopropyl Acetate	8.094	43	28694	2.98	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031921\
Data File : VN066283.D
Acq On : 19 Mar 2021 16:15
Operator : JC/MD
Sample : M1707-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ELEVATOR-511-EL-162-157

Quant Time: Mar 20 04:59:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

