

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031921\
 Data File : VN066284.D
 Acq On : 19 Mar 2021 16:38
 Operator : JC/MD
 Sample : M1712-24
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 YELLOW-WALL-COMP

Quant Time: Mar 20 05:00:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	307094	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	468749	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	432975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	158802	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	207223	46.84	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.68%
35) Dibromofluoromethane	7.507	113	170260	56.94	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.88%
50) Toluene-d8	10.028	98	629928	54.48	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.96%
62) 4-Bromofluorobenzene	12.345	95	175583	42.13	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.26%
Target Compounds						
					Qvalue	
16) Acetone	3.741	43	29805	14.02	ug/l	99
20) Methylene Chloride	4.457	84	15067	4.53	ug/l	94
43) Isopropyl Acetate	8.091	43	25392	2.42	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031921\
Data File : VN066284.D
Acq On : 19 Mar 2021 16:38
Operator : JC/MD
Sample : M1712-24
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
YELLOW-WALL-COMP

Quant Time: Mar 20 05:00:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

