

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031921\
 Data File : VN066288.D
 Acq On : 19 Mar 2021 18:11
 Operator : JC/MD
 Sample : M1707-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ELEVATOR-511-EL-167-162

Quant Time: Mar 20 05:01:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	299480	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	454222	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	433620	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	157439	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	212020	49.14	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 98.28%			
35) Dibromofluoromethane	7.507	113	167359	57.76	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 115.52%			
50) Toluene-d8	10.028	98	624634	55.75	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 111.50%			
62) 4-Bromofluorobenzene	12.345	95	177395	43.92	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 87.84%			
Target Compounds						Qvalue
16) Acetone	3.744	43	84481	40.76	ug/l	98
20) Methylene Chloride	4.455	84	20262	6.25	ug/l	91
43) Isopropyl Acetate	8.094	43	28217	2.78	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031921\
Data File : VN066288.D
Acq On : 19 Mar 2021 18:11
Operator : JC/MD
Sample : M1707-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ELEVATOR-511-EL-167-162

Quant Time: Mar 20 05:01:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

