

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031921\
 Data File : VN066290.D
 Acq On : 19 Mar 2021 18:58
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Manual Integrations
 APPROVED

MMDadoda
 3/22/2021 11:50:36 AM

Quant Time: Mar 20 05:02:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	284613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	421340	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	332227	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	120787	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	180717	44.07	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	88.14%		
35) Dibromofluoromethane	7.509	113	145208	54.02	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	108.04%		
50) Toluene-d8	10.028	98	506303	48.72	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.44%		
62) 4-Bromofluorobenzene	12.345	95	141084	37.66	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	75.32%#		
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	6.726	96	3268m	0.96	ug/l	
44) Trichloroethene	8.767	130	10763	3.46	ug/l	83
65) Chlorobenzene	11.371	112	5482	0.82	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031921\
Data File : VN066290.D
Acq On : 19 Mar 2021 18:58
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Manual Integrations
APPROVED

MMDadoda
3/22/2021 11:50:36 AM

Quant Time: Mar 20 05:02:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

