

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
 Data File : VN071389.D
 Acq On : 19 Mar 2022 11:30
 Operator : JC\MD
 Sample : VN0319WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0319WBL01

Quant Time: Mar 21 05:49:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	648693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1081153	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	982394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	324807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	420643	47.313	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.620%
35) Dibromofluoromethane	8.022	113	324081	47.502	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.000%
50) Toluene-d8	10.439	98	1293700	47.096	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.200%
62) 4-Bromofluorobenzene	12.727	95	457118	49.143	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
Data File : VN071389.D
Acq On : 19 Mar 2022 11:30
Operator : JC\MD
Sample : VN0319WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0319WBL01

Quant Time: Mar 21 05:49:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

