

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
 Data File : VN071402.D
 Acq On : 19 Mar 2022 16:55
 Operator : JC\MD
 Sample : N1958-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-9-8-9

Quant Time: Mar 21 05:53:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	768883	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1308459	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1253494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	427922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	498035	47.262	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.520%
35) Dibromofluoromethane	8.022	113	377617	45.734	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.460%
50) Toluene-d8	10.439	98	1652027	49.693	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.380%
62) 4-Bromofluorobenzene	12.727	95	579317	51.461	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.920%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	42553	8.299	ug/l	100
18) Methyl Acetate	4.851	43	20064	1.913	ug/l #	71
20) Methylene Chloride	5.104	84	17884	2.276	ug/l #	96
43) Isopropyl Acetate	8.557	43	75262	2.764	ug/l #	82
95) Naphthalene	15.504	128	17441	5.867	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
Data File : VN071402.D
Acq On : 19 Mar 2022 16:55
Operator : JC\MD
Sample : N1958-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-9-8-9

Quant Time: Mar 21 05:53:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

