

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
 Data File : VN071405.D
 Acq On : 19 Mar 2022 18:07
 Operator : JC\MD
 Sample : N1962-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-12-12-14

Quant Time: Mar 21 05:53:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	745415	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1262304	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1202211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	437530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	485836	47.556	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.120%
35) Dibromofluoromethane	8.016	113	367370	46.120	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.240%
50) Toluene-d8	10.439	98	1590803	49.601	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.200%
62) 4-Bromofluorobenzene	12.727	95	583557	53.733	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.460%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	89814	18.068	ug/l	98
18) Methyl Acetate	4.857	43	17670	1.781	ug/l	93
20) Methylene Chloride	5.104	84	10313	1.508	ug/l	96
43) Isopropyl Acetate	8.557	43	103455	3.939	ug/l #	81
95) Naphthalene	15.504	128	171319	12.743	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
Data File : VN071405.D
Acq On : 19 Mar 2022 18:07
Operator : JC\MD
Sample : N1962-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-12-12-14

Quant Time: Mar 21 05:53:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

