

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
 Data File : VN071407.D
 Acq On : 19 Mar 2022 18:56
 Operator : JC\MD
 Sample : N1962-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-18-22-25

Quant Time: Mar 21 05:54:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	738191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1262315	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1219393	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	442755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	487817	48.217	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.440%		
35) Dibromofluoromethane	8.022	113	367340	46.116	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.240%		
50) Toluene-d8	10.439	98	1599470	49.871	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.740%		
62) 4-Bromofluorobenzene	12.727	95	584232	53.795	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	107.580%		
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	83943	17.052	ug/l	91
18) Methyl Acetate	4.857	43	18728	1.873	ug/l	92
20) Methylene Chloride	5.104	84	10555	1.546	ug/l	87
40) Benzene	8.457	78	98929	2.573	ug/l	97
43) Isopropyl Acetate	8.557	43	99657	3.794	ug/l #	82
95) Naphthalene	15.504	128	88744	8.997	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
Data File : VN071407.D
Acq On : 19 Mar 2022 18:56
Operator : JC\MD
Sample : N1962-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-18-22-25

Quant Time: Mar 21 05:54:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

