

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
 Data File : VN071408.D
 Acq On : 19 Mar 2022 19:20
 Operator : JC\MD
 Sample : N1962-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-16-15-16

Quant Time: Mar 21 05:54:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	719421	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1220047	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1179142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	411516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	473716	48.045	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.080%
35) Dibromofluoromethane	8.016	113	358769	46.600	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.200%
50) Toluene-d8	10.439	98	1554526	50.149	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.300%
62) 4-Bromofluorobenzene	12.727	95	550822	52.476	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.960%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	90574	18.879	ug/l	99
18) Methyl Acetate	4.869	43	18707	1.908	ug/l	91
20) Methylene Chloride	5.092	84	9613	1.470	ug/l #	88
43) Isopropyl Acetate	8.563	43	110455	4.351	ug/l #	80
95) Naphthalene	15.504	128	99311	9.798	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
Data File : VN071408.D
Acq On : 19 Mar 2022 19:20
Operator : JC\MD
Sample : N1962-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-16-15-16

Quant Time: Mar 21 05:54:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

