

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
 Data File : VN071409.D
 Acq On : 19 Mar 2022 19:44
 Operator : JC\MD
 Sample : N1962-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-17-16-17

Quant Time: Mar 21 05:54:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	696084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1183496	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1143137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	402999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	468022	49.058	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.120%
35) Dibromofluoromethane	8.022	113	346680	46.420	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.840%
50) Toluene-d8	10.439	98	1507818	50.144	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.280%
62) 4-Bromofluorobenzene	12.727	95	548078	53.827	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.660%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	76181	16.412	ug/l	96
18) Methyl Acetate	4.863	43	18183	1.915	ug/l	92
20) Methylene Chloride	5.098	84	11712	1.752	ug/l #	79
43) Isopropyl Acetate	8.563	43	106362	4.319	ug/l #	81
95) Naphthalene	15.504	128	74769	8.705	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
Data File : VN071409.D
Acq On : 19 Mar 2022 19:44
Operator : JC\MD
Sample : N1962-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-17-16-17

Quant Time: Mar 21 05:54:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

