

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN031922\  
 Data File : VN071410.D  
 Acq On : 19 Mar 2022 20:08  
 Operator : JC\MD  
 Sample : N1962-15  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 C-18-17-18

Quant Time: Mar 21 05:55:10 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N030722W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Mar 08 06:52:46 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 8.087  | 168   | 773759   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.963  | 114   | 1322053  | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.739 | 117   | 1253408  | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.669 | 152   | 448134   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.434  | 65    | 500568   | 47.203   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =      | 94.400%  |
| 35) Dibromofluoromethane    | 8.016  | 113   | 380152   | 45.567   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =      | 91.140%  |
| 50) Toluene-d8              | 10.439 | 98    | 1648679  | 49.083   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =      | 98.160%  |
| 62) 4-Bromofluorobenzene    | 12.727 | 95    | 591041   | 51.963   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =      | 103.920% |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 4.287  | 43    | 71852    | 13.925   | ug/l   | 98       |
| 18) Methyl Acetate          | 4.863  | 43    | 19508    | 1.865    | ug/l # | 90       |
| 20) Methylene Chloride      | 5.099  | 84    | 11287    | 1.570    | ug/l   | 87       |
| 43) Isopropyl Acetate       | 8.557  | 43    | 101917   | 3.705    | ug/l # | 83       |
| 95) Naphthalene             | 15.504 | 128   | 98197    | 9.363    | ug/l   | 99       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN031922\  
Data File : VN071410.D  
Acq On : 19 Mar 2022 20:08  
Operator : JC\MD  
Sample : N1962-15  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
C-18-17-18

Quant Time: Mar 21 05:55:10 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N030722W.M  
Quant Title : SW846 8260  
QLast Update : Tue Mar 08 06:52:46 2022  
Response via : Initial Calibration

