

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
 Data File : VN071411.D
 Acq On : 19 Mar 2022 20:33
 Operator : JC\MD
 Sample : N1962-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-19-18-19

Quant Time: Mar 21 05:55:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	739396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1261706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1230827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	448827	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	492817	48.632	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.260%
35) Dibromofluoromethane	8.022	113	367539	46.163	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.320%
50) Toluene-d8	10.439	98	1594046	49.726	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.460%
62) 4-Bromofluorobenzene	12.727	95	583746	53.776	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.560%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	83834	17.002	ug/l	90
18) Methyl Acetate	4.869	43	18752	1.873	ug/l #	51
20) Methylene Chloride	5.093	84	10709	1.561	ug/l #	90
43) Isopropyl Acetate	8.563	43	101656	3.872	ug/l #	82
95) Naphthalene	15.504	128	141178	11.234	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031922\
Data File : VN071411.D
Acq On : 19 Mar 2022 20:33
Operator : JC\MD
Sample : N1962-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-19-18-19

Quant Time: Mar 21 05:55:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

