

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031924\
 Data File : VN081462.D
 Acq On : 19 Mar 2024 11:38
 Operator : JC\MD
 Sample : VN0319MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0319MBL01

Quant Time: Mar 20 01:14:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	453594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	766650	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	669244	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	280044	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	278900	48.215	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.420%
35) Dibromofluoromethane	8.165	113	238718	49.339	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.680%
50) Toluene-d8	10.565	98	896200	51.241	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.480%
62) 4-Bromofluorobenzene	12.847	95	282201	45.791	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031924\
Data File : VN081462.D
Acq On : 19 Mar 2024 11:38
Operator : JC\MD
Sample : VN0319MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0319MBL01

Quant Time: Mar 20 01:14:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

