

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031924\
 Data File : VN081466.D
 Acq On : 19 Mar 2024 13:23
 Operator : JC\MD
 Sample : P1769-02DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1DL

Quant Time: Mar 20 01:16:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	441370	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	752395	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	651440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	262060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	281244	49.967	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.940%
35) Dibromofluoromethane	8.165	113	230017	48.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.880%
50) Toluene-d8	10.565	98	877679	51.132	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.260%
62) 4-Bromofluorobenzene	12.853	95	276683	45.746	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.500%
Target Compounds						
					Qvalue	
52) Toluene	10.629	92	225089	15.159	ug/l	96
68) m/p-Xylenes	12.071	106	16859	1.681	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031924\
Data File : VN081466.D
Acq On : 19 Mar 2024 13:23
Operator : JC\MD
Sample : P1769-02DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1DL

Quant Time: Mar 20 01:16:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

