

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031924\
 Data File : VN081474.D
 Acq On : 19 Mar 2024 16:33
 Operator : JC\MD
 Sample : P1792-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11977

Quant Time: Mar 20 01:19:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	414723	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	743240	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	657487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	268293	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	276958	52.367	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.740%	
35) Dibromofluoromethane	8.165	113	231260	49.303	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.600%	
50) Toluene-d8	10.565	98	888489	52.400	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.800%	
62) 4-Bromofluorobenzene	12.847	95	275743	46.152	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 92.300%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.518	59	12309	13.355	ug/l	# 91
16) Acetone	4.430	43	45860	26.831	ug/l	100
20) Methylene Chloride	5.271	84	25500	4.555	ug/l	# 81
25) 2-Butanone	7.483	43	50192	17.121	ug/l	92
37) Ethyl Acetate	7.565	43	8290	1.226	ug/l	# 94
43) Isopropyl Acetate	8.688	43	437550	39.052	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031924\
Data File : VN081474.D
Acq On : 19 Mar 2024 16:33
Operator : JC\MD
Sample : P1792-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11977

Quant Time: Mar 20 01:19:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

