

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
 Data File : VN086011.D
 Acq On : 19 Mar 2025 16:02
 Operator : JC\MD
 Sample : PB167194ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167194ZHE#01

Quant Time: Mar 20 02:43:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	137571	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	239578	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	239167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	93887	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	102030	54.161	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.320%	
35) Dibromofluoromethane	8.159	113	85757	51.283	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.560%	
50) Toluene-d8	10.565	98	315967	52.062	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.120%	
62) 4-Bromofluorobenzene	12.847	95	101163	46.740	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 93.480%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	18142	31.523	ug/l	98
20) Methylene Chloride	5.271	84	3524	2.093	ug/l	95
25) 2-Butanone	7.483	43	13291	16.005	ug/l	99
43) Isopropyl Acetate	8.682	43	73746	18.320	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
Data File : VN086011.D
Acq On : 19 Mar 2025 16:02
Operator : JC\MD
Sample : PB167194ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167194ZHE#01

Quant Time: Mar 20 02:43:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

