

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
 Data File : VN086012.D
 Acq On : 19 Mar 2025 16:26
 Operator : JC\MD
 Sample : PB167194ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167194ZHE#02

Quant Time: Mar 20 02:45:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	138135	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	241700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	238319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	102950	54.427	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.860%	
35) Dibromofluoromethane	8.165	113	87881	52.092	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.180%	
50) Toluene-d8	10.565	98	318179	51.967	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.940%	
62) 4-Bromofluorobenzene	12.847	95	100278	45.924	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 91.840%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	17676	30.587	ug/l	90
20) Methylene Chloride	5.283	84	3446	2.038	ug/l #	82
25) 2-Butanone	7.488	43	13683	16.409	ug/l	97
43) Isopropyl Acetate	8.688	43	75469	18.584	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
Data File : VN086012.D
Acq On : 19 Mar 2025 16:26
Operator : JC\MD
Sample : PB167194ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167194ZHE#02

Quant Time: Mar 20 02:45:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

