

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
 Data File : VN086013.D
 Acq On : 19 Mar 2025 16:50
 Operator : JC\MD
 Sample : PB167194ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167194ZHE#03

Quant Time: Mar 20 02:48:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	136926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	241881	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	238409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	104159	55.552	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.100%			
35) Dibromofluoromethane	8.165	113	88699	52.537	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.080%			
50) Toluene-d8	10.565	98	320214	52.260	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.520%			
62) 4-Bromofluorobenzene	12.847	95	101256	46.337	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 92.680%			
Target Compounds						
16) Acetone	4.424	43	17253	30.119	ug/l	93
20) Methylene Chloride	5.271	84	3691	2.203	ug/l #	83
25) 2-Butanone	7.482	43	13298	16.089	ug/l	92
43) Isopropyl Acetate	8.688	43	78820	19.394	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
Data File : VN086013.D
Acq On : 19 Mar 2025 16:50
Operator : JC\MD
Sample : PB167194ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167194ZHE#03

Quant Time: Mar 20 02:48:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

