

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
 Data File : VN086014.D
 Acq On : 19 Mar 2025 17:14
 Operator : JC\MD
 Sample : PB167194ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167194ZHE#04

Quant Time: Mar 20 01:34:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	139880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	251910	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	243462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	105080	54.860	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.720%	
35) Dibromofluoromethane	8.165	113	88435	50.296	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.600%	
50) Toluene-d8	10.565	98	325939	51.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.160%	
62) 4-Bromofluorobenzene	12.847	95	100322	44.082	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 88.160%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	17639	30.143	ug/l	95
20) Methylene Chloride	5.271	84	3616	2.112	ug/l #	78
25) 2-Butanone	7.483	43	13854	16.407	ug/l	91
43) Isopropyl Acetate	8.688	43	81448	19.243	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
Data File : VN086014.D
Acq On : 19 Mar 2025 17:14
Operator : JC\MD
Sample : PB167194ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167194ZHE#04

Quant Time: Mar 20 01:34:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

