

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
 Data File : VN086016.D
 Acq On : 19 Mar 2025 18:03
 Operator : JC\MD
 Sample : Q1597-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1-CONCRETE-SLAB

Quant Time: Mar 20 01:35:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	140558	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	248810	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	239173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	104234	54.155	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.320%	
35) Dibromofluoromethane	8.165	113	88894	51.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.380%	
50) Toluene-d8	10.565	98	318109	50.470	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.940%	
62) 4-Bromofluorobenzene	12.847	95	106127	47.214	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 94.420%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	16659	28.331	ug/l	89
20) Methylene Chloride	5.277	84	3427	1.992	ug/l #	92
25) 2-Butanone	7.483	43	14744	17.377	ug/l #	84
43) Isopropyl Acetate	8.683	43	44800	10.716	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
Data File : VN086016.D
Acq On : 19 Mar 2025 18:03
Operator : JC\MD
Sample : Q1597-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1-CONCRETE-SLAB

Quant Time: Mar 20 01:35:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

