

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
 Data File : VN086017.D
 Acq On : 19 Mar 2025 18:27
 Operator : JC\MD
 Sample : Q1597-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2-CONCRETE-SLAB

Quant Time: Mar 20 06:34:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	138927	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	245130	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	236624	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	94379	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107313	56.410	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	112.820%	
35) Dibromofluoromethane	8.165	113	89976	52.588	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	105.180%	
50) Toluene-d8	10.565	98	304670	49.064	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	98.120%	
62) 4-Bromofluorobenzene	12.847	95	106618	48.144	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	96.280%	
Target Compounds						Qvalue
16) Acetone	4.436	43	17331	29.820	ug/l	99
20) Methylene Chloride	5.283	84	3431	2.018	ug/l #	79
25) 2-Butanone	7.483	43	14388	17.157	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
Data File : VN086017.D
Acq On : 19 Mar 2025 18:27
Operator : JC\MD
Sample : Q1597-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2-CONCRETE-SLAB

Quant Time: Mar 20 06:34:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

