

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
 Data File : VN086018.D
 Acq On : 19 Mar 2025 18:51
 Operator : JC\MD
 Sample : Q1597-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 3-CONCRETE-SLAB

Quant Time: Mar 20 01:36:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	257297	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	242349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	98553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	111480	56.221	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.440%	
35) Dibromofluoromethane	8.159	113	92212	51.346	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.700%	
50) Toluene-d8	10.565	98	314706	48.284	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.560%	
62) 4-Bromofluorobenzene	12.847	95	108879	46.840	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 93.680%	
Target Compounds						
						Qvalue
16) Acetone	4.424	43	17053	28.150	ug/l	93
20) Methylene Chloride	5.271	84	3510	1.981	ug/l	91
25) 2-Butanone	7.488	43	16139	18.463	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
Data File : VN086018.D
Acq On : 19 Mar 2025 18:51
Operator : JC\MD
Sample : Q1597-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
3-CONCRETE-SLAB

Quant Time: Mar 20 01:36:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

