

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
 Data File : VN086021.D
 Acq On : 19 Mar 2025 20:03
 Operator : JC\MD
 Sample : Q1604-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FRAC-TANK-N45878

Quant Time: Mar 20 01:38:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	136693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	249028	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	224312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	97536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	105468	56.346	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	112.700%	
35) Dibromofluoromethane	8.165	113	89064	51.240	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.480%	
50) Toluene-d8	10.565	98	299005	47.398	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.800%	
62) 4-Bromofluorobenzene	12.847	95	102473	45.548	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	91.100%	
Target Compounds						
16) Acetone	4.436	43	9205	16.097	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031925\
Data File : VN086021.D
Acq On : 19 Mar 2025 20:03
Operator : JC\MD
Sample : Q1604-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FRAC-TANK-N45878

Quant Time: Mar 20 01:38:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

