

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032019\
 Data File : VN054466.D
 Acq On : 20 Mar 2019 16:28
 Operator : JC/SP
 Sample : K1958-03
 Misc : 7.66g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Mar 21 10:00:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 14:42:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	613851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	955598	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	820057	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	335080	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	372359	48.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.54%	
35) Dibromofluoromethane	7.59	113	292088	47.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.88%	
50) Toluene-d8	10.09	98	1157340	50.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.02%	
62) 4-Bromofluorobenzene	12.40	95	372275	47.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.54%	
Target Compounds						
93) 1,2,4-Trichlorobenzene	14.91	180	11481	2.143	ug/l	91
95) Naphthalene	15.13	128	14454	1.209	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032019\
Data File : VN054466.D
Acq On : 20 Mar 2019 16:28
Operator : JC/SP
Sample : K1958-03
Misc : 7.66µL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Mar 21 10:00:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 14:42:09 2019
Response via : Initial Calibration

