

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032019\
 Data File : VN054468.D
 Acq On : 20 Mar 2019 17:23
 Operator : JC/SP
 Sample : K1958-01
 Misc : 6.79g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Mar 21 10:03:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 14:42:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	663721	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1002699	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	829125	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	324520	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	384760	46.61	ug/l	0.00
Spiked Amount			Recovery	=	93.22%	
35) Dibromofluoromethane	7.59	113	301377	46.65	ug/l	0.00
Spiked Amount			Recovery	=	93.30%	
50) Toluene-d8	10.09	98	1194443	49.18	ug/l	0.00
Spiked Amount			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	12.40	95	372157	45.03	ug/l	0.00
Spiked Amount			Recovery	=	90.06%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032019\
Data File : VN054468.D
Acq On : 20 Mar 2019 17:23
Operator : JC/SP
Sample : K1958-01
Misc : 6.79g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Mar 21 10:03:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 14:42:09 2019
Response via : Initial Calibration

