

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN032019\
 Data File : VN054473.D
 Acq On : 20 Mar 2019 19:40
 Operator : JC/SP
 Sample : K1994-05
 Misc : 4.34g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 40021

Quant Time: Mar 22 05:24:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	707003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1038049	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	856832	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	371718	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	399320	45.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.82%	
35) Dibromofluoromethane	7.59	113	308670	46.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.32%	
50) Toluene-d8	10.09	98	1239232	49.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.58%	
62) 4-Bromofluorobenzene	12.40	95	393878	46.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.08%	
Target Compounds						
					Qvalue	
68) m/p-Xylenes	11.62	106	3099	0.266	ug/l	89
69) o-Xylene	11.95	106	2976	0.261	ug/l	95
78) n-propylbenzene	12.59	91	18939	0.583	ug/l	97
80) 1,3,5-Trimethylbenzene	12.73	105	38407	1.564	ug/l	97
84) 1,2,4-Trimethylbenzene	13.04	105	92654	3.846	ug/l	96
85) sec-Butylbenzene	13.17	105	33629	1.262	ug/l	95
86) p-Isopropyltoluene	13.29	119	28661	1.203	ug/l	96
89) n-Butylbenzene	13.61	91	50753	2.723	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032019\
Data File : VN054473.D
Acq On : 20 Mar 2019 19:40
Operator : JC/SP
Sample : K1994-05
Misc : 4.34g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
40021

Quant Time: Mar 22 05:24:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

