

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032019\
 Data File : VN054483.D
 Acq On : 21 Mar 2019 1:10
 Operator : JC/SP
 Sample : PB118077TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB118077TB

Quant Time: Mar 22 05:49:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	723658	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1110981	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	928683	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	340428	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	426253	47.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.72%	
35) Dibromofluoromethane	7.59	113	341735	47.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.50%	
50) Toluene-d8	10.09	98	1312546	48.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.56%	
62) 4-Bromofluorobenzene	12.40	95	408399	44.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032019\
Data File : VN054483.D
Acq On : 21 Mar 2019 1:10
Operator : JC/SP
Sample : PB118077TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB118077TB

Quant Time: Mar 22 05:49:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

