

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032019\
 Data File : VN054494.D
 Acq On : 21 Mar 2019 6:12
 Operator : JC/SP
 Sample : K1943-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190129-LOC-6

Manual Integrations
 APPROVED

MMDadoda
 3/22/2019 3:51:35 PM

Quant Time: Mar 22 07:11:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	732520	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1114396	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	921122	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	333454	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	436652	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
35) Dibromofluoromethane	7.59	113	351794	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
50) Toluene-d8	10.09	98	1319521	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
62) 4-Bromofluorobenzene	12.40	95	400395	43.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.20%	
Target Compounds						
8) Diethyl Ether	3.41	74	4631	1.037	ug/l #	52
16) Acetone	3.83	43	22336m	8.154	ug/l	
20) Methylene Chloride	4.54	84	8175138	1012.397	ug/l #	85
43) Isopropyl Acetate	8.17	43	15754	1.259	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032019\
Data File : VN054494.D
Acq On : 21 Mar 2019 6:12
Operator : JC/SP
Sample : K1943-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20190129-LOC-6

Manual Integrations
APPROVED

MMDadoda
3/22/2019 3:51:35 PM

Quant Time: Mar 22 07:11:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

