

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032019\
 Data File : VN054496.D
 Acq On : 21 Mar 2019 7:07
 Operator : JC/SP
 Sample : K1943-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190127-LOC-4

Quant Time: Mar 22 07:15:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	730392	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1142092	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	927590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	345883	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	448361	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
35) Dibromofluoromethane	7.59	113	354469	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
50) Toluene-d8	10.09	98	1341355	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
62) 4-Bromofluorobenzene	12.40	95	407594	43.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.60%	
Target Compounds						
8) Diethyl Ether	3.42	74	5891	1.322	ug/l	93
16) Acetone	3.82	43	22862	8.370	ug/l #	82
20) Methylene Chloride	4.55	84	66681	8.282	ug/l	87
43) Isopropyl Acetate	8.17	43	17018	1.327	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032019\
Data File : VN054496.D
Acq On : 21 Mar 2019 7:07
Operator : JC/SP
Sample : K1943-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20190127-LOC-4

Quant Time: Mar 22 07:15:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

