

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032019\
 Data File : VN054498.D
 Acq On : 21 Mar 2019 8:02
 Operator : JC/SP
 Sample : K1943-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190131-LOC-8

Quant Time: Mar 22 07:19:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	713724	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1104405	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	895646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	329300	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	428948	48.32	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	96.64%	
35) Dibromofluoromethane	7.59	113	340664	47.88	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	95.76%	
50) Toluene-d8	10.09	98	1290672	48.25	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	96.50%	
62) 4-Bromofluorobenzene	12.40	95	391638	43.03	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	86.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.41	74	4378	1.006	ug/l	54
16) Acetone	3.82	43	20427	7.653	ug/l #	79
20) Methylene Chloride	4.55	84	181745	23.100	ug/l #	81
43) Isopropyl Acetate	8.17	43	16132	1.301	ug/l #	91
95) Naphthalene	15.13	128	44345	3.775	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032019\
 Data File : VN054498.D
 Acq On : 21 Mar 2019 8:02
 Operator : JC/SP
 Sample : K1943-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190131-LOC-8

Quant Time: Mar 22 07:19:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

