

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032019\
 Data File : VN054480.D
 Acq On : 20 Mar 2019 23:48
 Operator : JC/SP
 Sample : VN0320WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0320WBL01

Quant Time: Mar 22 05:44:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	721226	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1098729	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	918593	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	347915	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	424095	47.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.54%	
35) Dibromofluoromethane	7.59	113	330197	46.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.30%	
50) Toluene-d8	10.09	98	1299775	48.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.68%	
62) 4-Bromofluorobenzene	12.40	95	410349	45.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.64%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032019\
Data File : VN054480.D
Acq On : 20 Mar 2019 23:48
Operator : JC/SP
Sample : VN0320WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0320WBL01

Quant Time: Mar 22 05:44:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.00 min

Lab File: VN054480.D

Acq: 20 Mar 2019 23:48

Instrument :

MSVOA_N

ClientSampleId :


VN0320WBL01

Tot Ion:168 Resp: 721226

Ion Ratio Lower Upper

168 100

99 52.5 37.1 55.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.66



300000

200000

100000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 47.275 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. -0.00 min

Lab File: VN054480.D


Acq: 20 Mar 2019 23:48

Tgt Ion: 65 Resp: 424095

Ion Ratio Lower Upper

65 100

67 52.4 0.0 104.6



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0

8.03

200000


150000


100000

50000

0

Time-->









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.59 min Scan# 2119
Delta R.T. -0.00 min
Lab File: VN054480.D
Acq: 20 Mar 2019 23:48

Instrument :
MSVOA_N
ClientSampleId :
VN0320WBL01

Tot Ion:114 Resp: 1098729
Ion Ratio Lower Upper
114 100
63 20.9 0.0 35.2
88 16.0 0.0 30.2




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VN
Ion 87.90 (87.60 to 88.60): VN






#35
Dibromofluoromethane
Concen: 46.648 ug/l
RT: 7.59 min Scan# 1809
Delta R.T. -0.00 min
Lab File: VN054480.D
Acq: 20 Mar 2019 23:48

Tgt Ion:113 Resp: 330197
Ion Ratio Lower Upper
113 100
111 102.7 82.2 123.4
192 22.0 19.2 28.8



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.844 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN054480.D

Acq: 20 Mar 2019 23:48

Instrument :

MSVOA_N

ClientSampleId :


VN0320WBL01

Tot Ion: 98 Resp: 1299775

Ion Ratio Lower Upper

98 100

100 63.8 51.2 76.8



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN

10.09



600000

400000

200000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 45.316 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN054480.D

Acq: 20 Mar 2019 23:48


Tgt Ion: 95 Resp: 410349

Ion Ratio Lower Upper

95 100

174 86.4 0.0 191.0

176 82.9 0.0 187.4



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

12.40

300000

250000

200000


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN054480.D
Acq: 20 Mar 2019 23:48

Instrument :
MSVOA_N
ClientSampleId :
VN0320WBL01

Tot Ion:117 Resp: 918593
Ion Ratio Lower Upper
117 100
82 56.2 40.0 60.0
119 31.9 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN054480.D
Acq: 20 Mar 2019 23:48

Tgt Ion:152 Resp: 347915
Ion Ratio Lower Upper
152 100
115 57.3 27.2 81.6
150 157.2 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

