

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN032020\
 Data File : VN060653.D
 Acq On : 20 Mar 2020 15:25
 Operator : JC/MD
 Sample : PB127692ZHE#10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127692ZHE#10

Quant Time: Mar 21 04:55:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	180885	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	300433	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	275869	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	123597	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	118001	47.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.04%	
35) Dibromofluoromethane	7.57	113	88374	48.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.30%	
50) Toluene-d8	10.08	98	369303	49.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.54%	
62) 4-Bromofluorobenzene	12.40	95	134096	49.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.62%	

Target Compounds						Qvalue
20) Methylene Chloride	4.52	84	4200	1.94	ug/l	96
43) Isopropyl Acetate	8.14	43	50548	9.75	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN032020\
Data File : VN060653.D
Acq On : 20 Mar 2020 15:25
Operator : JC/MD
Sample : PB127692ZHE#10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127692ZHE#10

Quant Time: Mar 21 04:55:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

