

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN032020\
 Data File : VN060657.D
 Acq On : 20 Mar 2020 16:56
 Operator : JC/MD
 Sample : L1749-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BU-02-031920

Quant Time: Mar 21 00:32:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	198004	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	327884	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	312910	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147328	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130597	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
35) Dibromofluoromethane	7.57	113	98701	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
50) Toluene-d8	10.08	98	409608	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	12.40	95	156375	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
Target Compounds						
16) Acetone	3.78	43	294798	306.18	ug/l	Qvalue 100
20) Methylene Chloride	4.52	84	5056	2.13	ug/l	98
43) Isopropyl Acetate	8.14	43	33764	5.97	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN032020\
Data File : VN060657.D
Acq On : 20 Mar 2020 16:56
Operator : JC/MD
Sample : L1749-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BU-02-031920

Quant Time: Mar 21 00:32:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

