

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN032020\
 Data File : VN060659.D
 Acq On : 20 Mar 2020 17:41
 Operator : JC/MD
 Sample : L1762-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-031920

Quant Time: Mar 21 00:39:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	198917	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	328182	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	309930	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142243	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	129348	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
35) Dibromofluoromethane	7.57	113	98168	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
50) Toluene-d8	10.08	98	410277	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
62) 4-Bromofluorobenzene	12.40	95	152156	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.79	43	16108	16.65	ug/l	100
20) Methylene Chloride	4.52	84	5196	2.18	ug/l	97
43) Isopropyl Acetate	8.14	43	47466	8.38	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN032020\
Data File : VN060659.D
Acq On : 20 Mar 2020 17:41
Operator : JC/MD
Sample : L1762-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-05-031920

Quant Time: Mar 21 00:39:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

