

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032023\
 Data File : VN077066.D
 Acq On : 20 Mar 2023 14:17
 Operator : JC\MD
 Sample : 01904-04DL 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE137-20230313DL

Quant Time: Mar 21 05:36:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	417938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	740490	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	652575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	248242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	305492	45.761	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.520%
35) Dibromofluoromethane	8.172	113	237610	45.820	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.640%
50) Toluene-d8	10.572	98	901760	48.680	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.360%
62) 4-Bromofluorobenzene	12.848	95	289753	44.266	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.540%
Target Compounds						
						Qvalue
44) Trichloroethene	9.354	130	209903	37.201	ug/l	92
64) Tetrachloroethene	11.107	164	1035	0.236	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032023\
 Data File : VN077066.D
 Acq On : 20 Mar 2023 14:17
 Operator : JC\MD
 Sample : 01904-04DL 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampleId :
 RE137-20230313DL

Quant Time: Mar 21 05:36:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

