

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032023\
 Data File : VN077072.D
 Acq On : 20 Mar 2023 16:42
 Operator : JC\MD
 Sample : 01941-02DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT180D2-20230313DL

Quant Time: Mar 21 05:38:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	368561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	668664	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	600297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	223566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	280098	47.578	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.160%
35) Dibromofluoromethane	8.172	113	215222	45.961	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.920%
50) Toluene-d8	10.572	98	817971	48.900	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.800%
62) 4-Bromofluorobenzene	12.854	95	272770	46.147	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.300%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.378	101	14521	2.969	ug/l	95
12) 1,1-Dichloroethene	4.354	96	7409	1.656	ug/l #	70
27) cis-1,2-Dichloroethene	7.501	96	1349	0.243	ug/l	74
44) Trichloroethene	9.360	130	159445	31.294	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032023\
Data File : VN077072.D
Acq On : 20 Mar 2023 16:42
Operator : JC\MD
Sample : 01941-02DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT180D2-20230313DL

Quant Time: Mar 21 05:38:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

