

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032023\
 Data File : VN077073.D
 Acq On : 20 Mar 2023 17:06
 Operator : JC\MD
 Sample : 01941-09 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D2-20230314

Quant Time: Mar 21 05:38:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	368976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	677392	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	584988	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	217273	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	282340	47.905	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.820%
35) Dibromofluoromethane	8.172	113	212707	44.838	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.680%
50) Toluene-d8	10.572	98	813607	48.013	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.020%
62) 4-Bromofluorobenzene	12.854	95	265942	44.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.820%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.360	130	163533	31.683	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032023\
Data File : VN077073.D
Acq On : 20 Mar 2023 17:06
Operator : JC\MD
Sample : 01941-09 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D2-20230314

Quant Time: Mar 21 05:38:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

