

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032023\
 Data File : VN077077.D
 Acq On : 20 Mar 2023 18:43
 Operator : JC\MD
 Sample : 01941-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT180D1-20230313

Quant Time: Mar 21 05:40:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	372023	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	669073	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	584252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	215500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	277035	46.620	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.240%
35) Dibromofluoromethane	8.172	113	215911	46.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	10.572	98	818162	48.882	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.760%
62) 4-Bromofluorobenzene	12.854	95	274224	46.365	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.720%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.384	101	20149	4.081	ug/l	95
12) 1,1-Dichloroethene	4.354	96	10275	2.275	ug/l	96
27) cis-1,2-Dichloroethene	7.501	96	7780	1.388	ug/l	81
30) Chloroform	7.978	83	10455	1.039	ug/l	100
38) Carbon Tetrachloride	8.366	117	2491	0.349	ug/l	84
44) Trichloroethene	9.360	130	988493	193.891	ug/l	97
55) 1,1,2-Trichloroethane	11.025	97	4940	0.945	ug/l	94
95) Naphthalene	15.648	128	5610	2.289	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032023\
Data File : VN077077.D
Acq On : 20 Mar 2023 18:43
Operator : JC\MD
Sample : 01941-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT180D1-20230313

Quant Time: Mar 21 05:40:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

