

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032024\
 Data File : VN081488.D
 Acq On : 20 Mar 2024 12:50
 Operator : JC\MD
 Sample : VN0320WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0320WBL01

Quant Time: Mar 21 01:16:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	533451	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	896902	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	771363	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	307453	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	321509	47.260	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.520%
35) Dibromofluoromethane	8.165	113	268590	47.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.900%
50) Toluene-d8	10.565	98	1047620	51.200	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.400%
62) 4-Bromofluorobenzene	12.853	95	325798	45.188	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	90.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032024\
Data File : VN081488.D
Acq On : 20 Mar 2024 12:50
Operator : JC\MD
Sample : VN0320WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0320WBL01

Quant Time: Mar 21 01:16:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

