

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032024\
 Data File : VN081494.D
 Acq On : 20 Mar 2024 15:23
 Operator : JC\MD
 Sample : P1784-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 21 01:20:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	523252	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	884284	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	775743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	320620	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	319724	47.914	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.820%
35) Dibromofluoromethane	8.171	113	270746	48.514	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.020%
50) Toluene-d8	10.571	98	1037215	51.414	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.820%
62) 4-Bromofluorobenzene	12.853	95	338750	47.655	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	95.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032024\
Data File : VN081494.D
Acq On : 20 Mar 2024 15:23
Operator : JC\MD
Sample : P1784-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 21 01:20:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

